



## **EU Declaration of Conformity**

**Product** Laboratory Equipment **File Number** P225

**Manufacturer** Cole-Parmer Ltd

Beacon Road

Stone, Staffordshire

ST15 0SA

United Kingdom

## This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration

Thermal Cycler - PRO qPCR

(reference the attached list of catalogue numbers)

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

Low Voltage Directive2014/35/EUEMC Directive2014/30/EURoHS Directive2011/65/EC

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

IEC/EN 61010-1:2010 UL 61010-1:2012 CAN/CSA C22.2 No. 61010-1:2012	Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements.	
IEC/EN 61010-2-010:2014 UL 61010A-2-010:2002 CAN/CSA C22.2 No. 61010-2-010:2004	Particular requirements for laboratory equipment for the heating of materials.	
IEC/EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class A).	
CFR Part 15: 2013	Subpart B - Unintentional radiators. (Class A)	

## Signed for and on behalf of the above manufacturer

**Additional Information** Year of CE Marking: 2014

Place of Issue Stone, Staffordshire, UK

Date of Issue

Authorised Representative

O3 January 2017

Steve Marriott

**Title** Technical Director

**Signature** 

Range	Catalogue Number	Part Code	Description
PRO	93945-14	PRIMEPRO48	Prime Pro 48 Real Time PCR system